

Table 3. Specificity**Compound****Concentration (ng/mL)**

THC	
Cannabinol	>100,000
11-hydroxy- Δ^9 -THC	7,500
11-nor- Δ^9 -THC-9-COOH	250
11-nor- Δ^9 -THC-9-COOH	50
Δ^8 -THC	>100,000
Δ^9 -THC	>100,000

OPI	
Codeine	300
Hydrocodone	500
Hydromorphone	500
Lavofloxacin	100,000
Levophanol	5000
Meperidine	>100,000
Morphine	300
Morphine-3- β -D-glucuronide	300
Nalorphine	15,000
Naloxone	>100,000
Norcodeine	>100,000
Oxycodone	5,000
Oxydone	20,000
Oxymorphone	10,000
Thebaine	>100,000
Tramadol	

CO C	
Benzoylegonine	300
Cocaine HCl	>100,000
Egonine HCl	>100,000

MET	
D-Amphetamine	>100,000
D,L-Amphetamine	>100,000
(-)Ephedrine	>100,000
(+)Ephedrine	>100,000
Isomethcptene	12500
D-Methamphetamine	1,000
p-OH-Methamphetamine	3,000
Methylenedioxymphetamine	>100,000
Methylenedioxymethylamphetamine(MDEA)	100,000
Methylenedioxymethamphetamine	1,000

The following compounds show no cross-reactivity when tested with AccuSign® DOA 4 at a concentration of 100 µg/mL. (Table 4.)

Table 4. Non Cross-Reacting Compounds

Acetaminophen	Chlordiazepoxide	Guaiacol glycerol ether
Acetylsalicylate	Chlorpheniramine	Hydrochlorothiazide
Aminopyrine	Chlorpromazine	Imipramine
Amitriptyline	Chloroquine	Lidocaine
Amobarbital	Dextropropoxyphene	Diazepam
Amoxapine	Diphenylhydantoin	Methadone
Ampicillin	Diphenhydramine	Methaqualone
Apomorphine	Epinephrine	Methyprylon
Ascorbic acid	Erythromycin	Naproxen
Atropine	Estreptokinase	Norethindrone
Benzocaine	Gentisic acid	Penicillin
Butabarbital	Glutethimide	Pentobarbital

AccuSign® is a Registered Trademark of Princeton BioMeditech Corporation.

Patent No.: 5,559,041



Phencyclidine
Phenolbutazone
Phenylpropanolamine
Prednisone

Secobarbital
Tetracycline
Tetrahydrozoline
Trifluoperazine

Tryptamine
Zomepirac

References

- Hawks RL, Chiang CN, eds. *Urine Testing for Drugs of Abuse*. National Institute on Drug Abuse (NIDA), Research Monograph 73; 1986.
- Tietz, Norbert W. *Textbook of Clinical Chemistry*. W.B. Saunders Company. 1986.
- Ambre J. *J. Anal. Toxicol.* 1985;9:241–5.
- Baselt RC. *Disposition of Toxic Drugs and Chemicals in Man*. 2nd Ed., Davis, CA: Biomedical Publ.; 1982.

Symbols Key

	Manufactured by
	CE Mark
	Authorized Representative
	In Vitro Diagnostic Medical Device
	Catalog Number
	Consult Instructions for Use
	Batch Code
	“Use By” date in year-month-day format
	Temperature Limitation
	Contains sufficient for <n> tests
	Do not reuse
	Contents
	Test Device
	Transfer Pipette
	Instructions for Use
	One-step immunochromatographic Assay for the Detection of Drugs of Abuse in Urine
	Marijuana/Opiates/Cocaine/Methamphetamine Test

P-5844-F

AccuSign® DOA 4

THC/OPI/COC/MET

New One-Step Panel Assay for Drugs of Abuse

For *In Vitro* Use Only

Simple One-Step Immunoassay for the Qualitative Detection of THC, Opiates, Cocaine, Methamphetamine, and/or their Metabolites in Urine

PBM

Catalog No. DOA-241-35 35 Test Kit
DOA-241-10 10 Test Kit

Intended Use

The AccuSign® DOA 4 THC/OPI/COC/MET Panel Assay is a simple, one-step immunochromatographic test for the rapid, qualitative detection of THC metabolite, opiates, cocaine metabolite, and methamphetamine in urine. The test detects the major metabolites of these drugs at the following cutoff concentrations.

THC	11-nor- Δ^9 -THC-9-carboxylic acid	50 ng/mL
OPI	Morphine	300 ng/mL
CO C	Benzoylegonine	300 ng/mL
MET	D-Methamphetamine	1000 ng/mL

The AccuSign® DOA 4 THC/OPI/COC/MET test provides only a preliminary analytical result. A more specific alternative chemical method must be used in order to obtain a confirmed analytical result. Gas chromatography, mass spectrometry (GC/MS) is the preferred confirmatory method. Other chemical confirmatory methods are available. Clinical consideration and professional judgment should be applied to any drug of abuse test result, particularly when preliminary positive results are used.¹

Summary and Explanation

THC (Δ^9 -tetrahydrocannabinol) is the primary active ingredient in cannabinoids (marijuana). When ingested or smoked, it produces euphoric effects. Users experience impairment of short term memory and THC use slows learning. Also, it may cause transient episodes of confusion, anxiety, or frank toxic delirium. Long term, relatively heavy use may be associated with behavioral disorders. The peak effect of smoking THC occurs in 20–30 minutes and the duration is 90–120 minutes after one cigarette. Elevated levels of urinary metabolites are found within hours of exposure and remain detectable for 3–10 days after smoking. The main metabolite excreted in the urine is 11-nor- Δ^9 -tetrahydrocannabinol-9-carboxylic acid.²

Morphine, codeine, and semisynthetic derivatives of morphine belong to the class of drugs called opiates. An opiate exerts its effects on the central nervous system and can produce euphoria, respiratory depression and coma when it is abused. Morphine is the prototype compound of opiates. Morphine is excreted in the urine as morphine-3-glucuronide, unchanged morphine, and other minor metabolites. Heroin is metabolized to morphine and codeine and excreted in the urine with a small amount in unchanged form. Codeine is also excreted as morphine and in the form of conjugates. Although some opiate metabolites appear in the feces, urinary excretion is the primary route of elimination.^{1,2,4}

Cocaine, derived from the leaves of coca plant, is a potent central nervous system (CNS) stimulant and a local anesthetic. Cocaine induces euphoria, confidence and a sense of increased energy in the user; these psychological effects are accompanied by increased heart rate, dilation of the pupils, fever, tremors and sweating. Cocaine is used by smoking, intravenous, intranasal or oral administration, and excreted in the urine primarily as benzoylecgonine in a short time. Benzoylecgonine has a longer biological half-life (5–8 hours) than cocaine (0.5–1.5 hours) and can generally be detected for 24–60 hours after cocaine use or exposure.^{3,4}

Methamphetamine is a potent sympathomimetic agent with therapeutic applications. The drug can be taken orally, injected, or inhaled. Acute higher doses lead to enhanced stimulation of the central nervous system and induce euphoria, alertness, reduced appetite, and a sense of increased energy and power. Cardiovascular responses to methamphetamine include increased blood pressure and cardiac arrhythmias. More acute responses include anxiety, paranoia, hallucinations, psychotic behavior, and eventually, depression and exhaustion. The effects of methamphetamine generally last 2–4 hours, and the drug has a half-life of 9–24 hours in the body. Methamphetamine is excreted in the urine primarily as amphetamine and oxidized and deaminated derivatives. However, 10–20% of methamphetamine is excreted unchanged. Thus, the presence of the parent compound in the urine indicates methamphetamine use. Methamphetamine is generally detectable in the urine for 3–5 days, depending on urine pH level.⁴

Principle


The AccuSign® DOA 4 test uses solid-phase chromatographic membrane immunoassay technology for the qualitative, simultaneous detection of THC, opiates, cocaine, and methamphetamine in human urine. The test is based on the principle of the highly specific immunochemical reactions between antigens and antibodies which are used for the analysis of specific substances in biological fluids. The test relies on the competition between the drug conjugates and the drugs which may be present in the urine sample, for binding to antibodies. In the test procedure, a sample of urine is placed in the Sample well of the device and is allowed to migrate upward. If the drug is present in the urine sample, it competes with the drug conjugate bound to the dye, for the limited antibodies immobilized on the membrane. If the level of drug or drug metabolite is above the cutoff level, the drug will saturate the antibodies, thus inhibiting the binding of the dye coated with drug conjugates to the antibodies on the membrane. This prevents the formation of a line on the membrane. Therefore, a drug-positive urine sample will not generate a line at the specific drug position in the Result window, indicating a positive result. A negative urine sample will generate a line at the specific drug position in the Result window, indicating a negative result. The same principle of competition is applicable where the drug conjugate is immobilized on the membrane and the antibody is coated on the dye.

In addition to the Test line(s) that may appear in the Result window, a Control line is present to confirm the viability of the test. This Control line (validation line) should always appear if the test is conducted properly. Polyclonal sheep anti-mouse IgG antibody is immobilized on the control line. The monoclonal antibody-dye conjugates that pass the line will be captured and produce a colored line at the Control position (C). This works as a procedural control, confirming that proper sample volume was used and the reagent system at the Control line and the conjugate-color indicator worked properly. If insufficient sample volume is used, there may not be a Control line, indicating the test is invalid.

Materials Provided

The AccuSign® DOA 4 test kit contains all the reagents necessary to perform the assay.

- AccuSign® DOA 4 device. The test device contains a membrane strip and a dye pad. The membrane strip is coated with THC-protein (a purified bovine protein) conjugate, monoclonal anti-morphine and anti-benzoylecgonine, and anti-methamphetamine antibodies. Sheep anti-mouse IgG antibody is coated for the control band. The dye pad contains colloidal gold coated with monoclonal anti-THC antibody, and conjugates of methamphetamine, morphine, and benzoylecgonine (each drug is conjugated with a purified bovine protein).
- Disposable sample dispenser.
- Instructions for use.



Precautions

- For *in vitro* diagnostic use only.
- Avoid cross contamination of urine samples by using a new urine specimen container and dropper for each urine sample.
- This test kit does not contain any HIV or hepatitis infective components.
- Urine specimens are potentially infectious. Proper handling and disposal methods should be followed, according to good laboratory practices.
- The AccuSign® device should remain in its original sealed pouch until ready for use.
- Do not use the test kit after the expiration date.

Storage and Stability

The AccuSign® DOA 4 test kit should be stored at 2–30°C (35–86°F) in the original sealed pouch. The expiration dating was established under these storage conditions.

Specimen Collection and Preparation

Approximately 110 μL of urine sample is required for each test. Fresh urine specimens do not require any special handling or pretreatment. Specimens should be collected in a clean glass or plastic container. If testing will not be performed immediately, specimens should be refrigerated (2–8°C) or frozen. Specimens should be brought to room temperature before testing.

Specimens containing a large amount of particulate matter may give inconsistent test results. These specimens should be clarified by centrifuging or allowing to settle before testing.

Test Procedure

The test procedure consists of adding the urine sample to the Sample well of the device and watching for the appearance of colored lines in the result window.

Test Protocol

- For each test, open one AccuSign® DOA 4 pouch and label the AccuSign® DOA 4 device with the patient ID.
- Holding the dropper vertically, dispense 3 full drops (110 μL) of the urine sample into the Sample well.
- Read the result after 5 minutes, but within 10 minutes.

Interpretation of Results

Negative: The appearance of a reddish-purple Control line (C) and a line for a specific drug indicates a negative test result; i.e., no drug above the cutoff level has been detected. The color intensities of the Control line and a specific drug line may not be equal. *Any faint line at a specific drug name in the Result window, visible in 10 minutes, should be interpreted as negative. A negative test result does not indicate the absence of drug in the sample; it only indicates the sample does not contain drug above the cutoff level in qualitative terms.*

Positive: The appearance of a reddish-purple Control line and no distinct line at a specific drug name indicates the test result is positive for that drug (i.e., the specimen contains the drug at a concentration above the cutoff level). *A positive test result does not provide any indication of the level of intoxication or urinary concentration of the drug in the sample; it only indicates the sample contains drug above the cutoff level in qualitative terms.*

Invalid: A distinct Control line (C) should always appear. The test is invalid if no Control line forms at the C position. Such tests should be repeated with a new AccuSign® DOA 4 test device.

Examples of possible results are shown in the diagram above.

- THC (-), Opiates (-), Cocaine (-), Methamphetamine (-):** Five reddish-purple lines—one Control line at the C position and one each at the THC, OPI, COC, and MET positions.
- THC (-), Opiates (-), Cocaine (-), Methamphetamine (+):** Four reddish-purple lines—one Control line at the C position and one line each at the THC, OPI, and COC positions; no line at the MET position.
- THC (+), Opiates (-), Cocaine (+), Methamphetamine (-):** Three reddish-purple lines—one Control line at the C position, one line each at the OPI and MET positions; no lines at the THC and COC positions.
- THC (-), Opiates (+), Cocaine (-), Methamphetamine (-):** Four reddish-purple lines—one Control line at the C position and one line each at the THC, COC, and MET positions; no line at the OPI position.
- Invalid:** No line at the C position.

There are other possible results, depending on the combinations of drugs present in the urine sample.

Limitations

- The test is designed for use with unadulterated urine only.
- There is a possibility that factors such as technical or procedural errors, as well as other substances in the urine sample than those listed in Table 4 below, may interfere with the test and cause erroneous results.
- Adulterants, such as bleach and/or alum, in urine specimens may produce

erroneous results regardless of the method of analysis. If adulteration is suspected, the test should be repeated with a new sample.

- The test result read after 10 minutes may not be consistent with the original reading obtained within the 10 minute reading period. The test result must be read within 10 minutes of sample application.
- Certain medications containing opiates or opiate derivatives or methamphetamines may produce a positive result. Additionally, foods and tea containing poppy products and/or coca leaves may produce a positive result. Prolonged passive smoking of THC may also produce a positive result.

User Quality Control

Internal Control: Each AccuSign® DOA 4 test device has a built-in control. The Control line is an internal positive procedural control. A distinct reddish-purple Control line should appear in the Control position, if the test procedure is performed properly, an adequate sample volume is used, the sample and reagent are wicking on the membrane, and the test reagents at the control line and the conjugate-color indicator are reactive. In addition, if the test is performed correctly and the device is working properly, the background in the Result window will become clear and provide a distinct result. This may be considered an internal negative procedural control.

The positive and negative procedural controls contained in each AccuSign® DOA 4 test device satisfy the requirements of testing a positive control and a negative control on a daily basis. If the Control line does not appear in the Control position, the test is invalid and a new test should be performed. If the problem persists, contact PBM for technical assistance.

External Control: External controls may also be used to assure that the reagents are working properly and that the assay procedure is followed correctly. It is recommended that a control be tested at regular intervals as good laboratory testing practice. For information on how to obtain controls, contact PBM's Technical Services.

Expected Values

AccuSign® DOA 4 is a qualitative assay. The amount of drugs and metabolites present in urine cannot be estimated by the assay. The assay results distinguish positive from negative samples. Positive results indicate the samples contain the specific drug above the cutoff concentration.

Performance Characteristics

The AccuSign® DOA 4 Panel Assay detects THC metabolite, opiates, cocaine metabolite, methamphetamine at cutoff levels based on the recommendations of the SAMHSA for screening of these drugs in urine.

THC	11-nor-Δ ⁹ -THC-9-carboxylic acid	50 ng/mL
OPI	Morphine	300 ng/mL
COC	Benzoylegonine	300 ng/mL
MET	D-Methamphetamine	1000 ng/mL

The accuracy of AccuSign® DOA 4 was evaluated in comparison to a commercially available immunoassay (Syva® EMIT® II) for each of these four drugs. The results are shown in Table 1.

Table 1. Accuracy: Comparison of AccuSign® DOA 4 with Syva® EMIT® II Assay

	Syva® EMIT® II (THC)			TOTAL
	Positive	Negative		
AccuSign®	Positive	327	5	332
DOA 4 (THC)	Negative	13	655	668
TOTAL		340	660	1000
Syva® EMIT® II (OPI)				
	Syva® EMIT® II (OPI)			TOTAL
	Positive	Negative		
AccuSign®	Positive	249	0	249
DOA 4 (OPI)	Negative	1	716	717
TOTAL		250	716	966

		Syva® EMIT® II (COC)		
		Positive	Negative	TOTAL
AccuSign®	Positive	369	3	372
DOA 4 (COC)	Negative	16	635	651
TOTAL		385	638	1023
		Syva® EMIT® II (AMP/MET)		
		Positive	Negative	TOTAL
AccuSign®	Positive	108	0	108
DOA 4 (MET)	Negative	28	184	212
TOTAL		136	184	320
		Relative Sensitivity	Relative Specificity	
THC	96.2% (327/340)	99.2% (655/660)		
Opiates	99.6% (249/250)	>99% (716/716)		
Cocaine	95.8% (369/385)	99.5% (635/638)		
Methamphetamine	79.4% (108/136)	>99% (184/184)		

In a separate study, AccuSign® DOA 4 was evaluated against specimens confirmed as positive by GC/MS, for each of the four drugs. The results demonstrate the excellent correlation of AccuSign® DOA 4 with GC/MS (99% agreement, Table 2).

Table 2. Accuracy: Comparison of AccuSign® DOA 4 with GC/MS Assay

	AccuSign®	GC/MS
THC	Positive	87
	Negative	1
OPI	Positive	73
	Negative	1
COC	Positive	77
	Negative	1
MET	Positive	88
	Negative	1

Precision and Accuracy

The precision of the AccuSign® DOA 4 Panel Assay was determined by carrying out the test with serially diluted standard drug solutions. About 98% of the samples containing 25% over the cutoff level of cocaine, opiates, methamphetamine and about 90% of the samples containing THC concentration 25% over the cutoff level consistently showed positive results.

The study also included over 40 samples ± 25% cutoff level as a challenge of cutoff precision. These results were found to be consistently in agreement with expected test results.

Distribution of Random Error:

Twenty (20) blind samples prepared by spiking various concentrations of cocaine, THC, morphine, or methamphetamine were separately tested by two operators. The test results from the two operators showed complete agreement.

Reproducibility

The reproducibility of the test results of the AccuSign® DOA 4 Panel Assay was examined at three different sites using a total of 15 blind controls, consisting of 5 negative samples, 5 moderately positive samples, and 5 strongly positive samples (i.e., a concentration 3 times the cutoff level). The results obtained at these three sites with these controls demonstrated 100% agreement with each other.

Specificity

The following table lists compounds that are detected by the AccuSign® DOA 4 test. The specificity of the AccuSign® DOA 4 test was determined by adding various drugs and drug metabolites to drug-negative urine specimens and testing with the AccuSign® DOA 4 test kit. The results are expressed in terms of the concentration required to produce a positive result (Table 3).